

Towards a reliable, non-invasive melanin assessment for pigmented skin

Spectrophotometry is used to estimate melanin density (MD) in Caucasians¹ but applicability of the calculation to people with pigmented skin, who also experience sun sensitivity, is not known. Here, we directly compared biopsy melanin concentration (MC) with Melanin Index (MI), Individual Typology Angle (ITA) values,² calculated Melanin Density (MD) and Self-Reported Sun Sensitivity (SRSS), with an aim to identify a non-invasive, reliable melanin assessment technique for deeply pigmented skin.

Participants (n = 50) were drawn from employees of the Council for Scientific and Industrial Research (CSIR) in Pretoria, South Africa

from May 9 to 17, 2016. The CSIR Research Ethics Committee approved the study protocol (Certificate number 79/2013). Healthy study participants gave written informed consent, spoke English, cleaned their non-dominant arm with a sanitary wipe, and answered a short questionnaire to self-identify sex, SRSS (just burn and not tan; burn first then tan afterwards; not burn at all, just tan) and population group (Black African; Indian/Asian; coloured (which is defined in South Africa as mixed population group); or White).³

We determined MI and ITA using a Mexameter MX 18 and Skin Colorimeter CL 400, respectively (both Courage + Khazaka




FIGURE 1 Distribution of participants' (A) MI values (n = 47), (B) ITA values (n = 49), (C) spectrophotometric MD (n = 49), and (D) MC (n = 49), vs SRSS (1 = just burn and not tan; 2 = burn first then tan afterwards; 3 = not burn at all, just tan). The upper whisker is the 95th percentile, the upper box line is the 75th percentile, the middle line in the box is the median, the lower line of box is the 25th percentile and the lower whisker is the 5th percentile [Colour figure can be viewed at wileyonlinelibrary.com]

	MC (n = 49) (21; 13-31)	MI (n = 47) (580; 294-868)	ITA (n = 49) (-17; -40-16)	MD (n = 49) (4; 3-5)	SRSS (n = 49) (1:7; 2:16; 3:26)
MC	1	0.331 (P = .02)	-0.276 (P = .05)	-0.280 (P = .05)	0.129 (P = .38)
MI		1	-0.864 (P < .001)	-0.783 (P < .001)	0.070 (P = .64)
ITA			1	0.921 (P < .001)	0.014 (P = .92)
MD				1	-0.059 (P = .69)

TABLE 1 Correlations of estimates of MC with MI, ITA, MD, and SRSS. Mean and range are given in brackets for ITA, MI, MD, and MC, with n per SRSS category provided for SRSS

Electronic) and MD¹ using a portable Spectrophotometer CM-2600d (Konica Minolta), taking an average of 3 measurements of skin on the inner surface of the upper, non-dominant arm. After a half-hour rest, a trained nurse took a 2-mm punch biopsy of the measurement area. Biopsies were fixed, stained, and analysed according to a standard protocol⁴ to determine MC. We used Stata (v.14.0) for statistical analysis, with Pearson's correlation coefficient (*r*) used to assess the correlation between MC and all other variables,¹ and linear regression to test predictive models of MC using likelihood ratios.

Overall, MC was available for 49 participants. Of these, 65% were women (with missing sex data for one participant), and all but 1 person (who self-reported as Coloured) self-identified as Black African. Participants did not relate to SRSS terms such as "sunburn" and "tan," evident in the lack of meaningful comparisons with objective measures of skin sensitivity to the sun (Figure 1).

There were statistically significant inverse correlations between MC and ITA, and MC and MD (Table 1) and there was a significant positive correlation between MC and MI but no significant correlation between MC and SRSS. Correlations between the other measures of skin type were as we have previously reported.^{5,6}

For the sample with full data, the best-fitting multiple linear regression model for MC (explaining 26% of the variance) included only sex ($\beta = -3.31$, $P = .004$) and MI ($\beta = 0.005$, $P = .23$). We tested different ways of combining the MD input data (reflectance at wavelengths of 400 and 420 nm), and adding phenotypic characteristics, but were unable to find any algorithm that explained more than 29% of the variance in biopsy MC ($P = .02$), which is much less than that previously shown for Caucasian skin (ie 68%).¹

Here, the best skin type predictor of MC was MI; however, MI alone explained only 11% of the variance in MC, with a greater contribution when sex was also included in the model. The spectrophotometric MD, previously shown to predict MC well in Caucasians,¹ was only a weak (and non-significant) predictor of MC in people with deeply pigmented skin. In studies of the effect of sun exposure on human health, for example, for production of vitamin D or for sun sensitivity to sunburn or immune suppression, an accurate measure for "skin type" is essential.

Although calculated MD is commonly used for this purpose, we show here that the algorithm developed for Caucasian populations is not suitable for use in people with deeply pigmented skin. Developing an accurate, non-invasive estimate of MC in people with deeply pigmented skin—or that is generalizable to the range of skin types—will prove useful in epidemiology studies, in clinical dermatology and application of laser therapy treatments.

ACKNOWLEDGEMENTS

We thank Dr Kevin Sevier for training our research nurse, Louise Renton for performing the laboratory work, Patricia Albers and Mirriam Mogotsi for assisting with data collection, and the participants in the study.

FUNDING

This study was supported in part by funding from the South African Medical Research Council and the National Research Foundation of South Africa.

CONFLICT OF INTEREST

The authors state no conflict of interest.

ORCID

C. Y. Wright  <http://orcid.org/0000-0001-9608-818X>

R. M. Lucas  <http://orcid.org/0000-0003-2736-3541>

J. L. du Plessis  <http://orcid.org/0000-0001-5122-8492>

T. Kapwata  <http://orcid.org/0000-0003-2518-6764>


Z. Kunene  <http://orcid.org/0000-0001-5655-5556>

Keywords

melanin, pigmented skin, skin biopsy, sun sensitivity

C. Y. Wright^{1,2} 

R. M. Lucas^{3,4} 

T. Kapwata⁵ 

Z. Kunene⁵ 

J. L. du Plessis⁶ 

¹Environment and Health Research Unit, South African Medical Research Council, Pretoria, South Africa

²Department of Geography, Geoinformatics and Meteorology, University of Pretoria, Pretoria, South Africa

³Research School of Population Health, National Centre for Epidemiology and Population Health, Australian National University, Canberra, Australia

⁴Centre for Ophthalmology and Visual Sciences, University of Western Australia, Perth, Australia

⁵Environment and Health Research Unit, South African Medical Research Council, Johannesburg, South Africa

⁶Occupational Hygiene and Health Research Initiative, North-West University, Potchefstroom, South Africa

Correspondence

C. Y. Wright, Environment and Health Research Unit, South African Medical Research Council, Pretoria, South Africa.

Email: cwright@mrc.ac.za

REFERENCES

1. Dwyer T, Muller HK, Blizzard L, Ashbolt R, Phillips G. The use of spectrophotometry to estimate melanin density in Caucasians. *Cancer Epidemiol Biomarkers Prev*. 1998;7:203-206.
2. Del Bino S, Bernerd F. Variations in skin colour and the biological consequences of ultraviolet radiation exposure. *Br J Dermatol*. 2013;169:33-40.

3. Statistics South Africa. Mid-year population estimates. 2017. <http://www.statssa.gov.za/publications/P0302/P03022017.pdf>. Accessed April 3, 2018.
4. Hacker E, Boyce Z, Kimlin MG, et al. The effect of MC1R variants and sunscreen on the response of human melanocytes in vivo to ultraviolet radiation and implications for melanoma. *Pigment Cell Melanoma Res.* 2013;23:835-844.
5. Wright CY, Karsten A, Wilkes M, et al. Diffuse reflectance spectroscopy versus Mexameter[®] MX18 measurements of melanin and erythema in an African population. *Photochem Photobiol.* 2016;92:632-636.
6. Wilkes M, Wright CY, du Plessis JL, Reeder AI. Fitzpatrick skin type, Individual Typology Angle and melanin index in an African population: taking steps toward universally applicable skin photosensitivity assessments. *JAMA Dermatol.* 2015;151:902-903.