



Correction for a measurement artifact of the Multi-Angle Absorption Photometer (MAAP) at high black carbon mass concentration levels

A.-P. Hyvärinen¹, V. Vakkari², L. Laakso^{1,3}, R. K. Hooda^{1,4}, V. P. Sharma⁴, T. S. Panwar^{4,5}, J. P. Beukes³, P. G. van Zyl³, M. Josipovic³, R. M. Garland^{6,7}, M. O. Andreae⁶, U. Pöschl⁶, and A. Petzold⁸

¹Finnish Meteorological Institute, P.O. Box 503, 00101 Helsinki, Finland

²Department of Physics, University of Helsinki, P.O. BOX 64, 00014 Helsinki, Finland

³School of Physical and Chemical Sciences, North-West University, Potchefstroom, South Africa

⁴The Energy and Resources Institute (TERI), Darbari Seth Block, IHC Complex, Lodhi Road, 110 003 New Delhi, India

⁵WWF India, Lodhi Road, 110 003 New Delhi, India

⁶Max Planck Institute for Chemistry, P.O. Box 3060, 55020 Mainz, Germany

⁷Natural Resources and the Environment, The Council for Scientific and Industrial Research (CSIR), Pretoria, South Africa

⁸Forschungszentrum Jülich GmbH, Institute of Energy and Climate Research IEK-8: Troposphere, 52425 Jülich, Germany

Correspondence to: A.-P. Hyvärinen (antti.hyvarinen@fmi.fi)

Received: 10 August 2012 – Published in Atmos. Meas. Tech. Discuss.: 12 September 2012

Revised: 13 December 2012 – Accepted: 14 December 2012 – Published: 11 January 2013

Abstract. The Multi-Angle Absorption Photometer (MAAP) is a widely-used instrument for aerosol black carbon (BC) measurements. In this paper, we show correction methods for an artifact found to affect the instrument accuracy in environments characterized by high black carbon concentrations. The artifact occurs after a filter spot change – as BC mass is accumulated on a fresh filter spot, the attenuation of the light (raw signal) is weaker than anticipated. This causes a sudden decrease, followed by a gradual increase in measured BC concentration. The artifact is present in the data when the BC concentration exceeds $\sim 3 \mu\text{g m}^{-3}$ at the typical MAAP flow rate of 16.7 L min^{-1} or $1 \text{ m}^3 \text{ h}^{-1}$. The artifact is caused by erroneous dark counts in the photodetector measuring the transmitted light, in combination with an instrument internal averaging procedure of the photodetector raw signals. It was found that, in addition to the erroneous temporal response of the data, concentrations higher than $9 \mu\text{g m}^{-3}$ (at the flow rate of 16.7 L min^{-1}) are underestimated by the MAAP. The underestimation increases with increasing BC accumulation rate. At a flow rate of 16.7 L min^{-1} and concentration of about $24 \mu\text{g m}^{-3}$ (BC accumulation rate $\sim 0.4 \mu\text{g min}^{-1}$), the underestimation is about 30 %. There are two ways of overcoming the MAAP artifact. One method is by logging the raw

signal of the 165° photomultiplier measuring the reflected light from the filter spot. As this signal is not affected by the artifact, it can be converted to approximately correct absorption and BC values. However, as the typical print formats of the MAAP do not give the reflected signal as an output, a semi-empirical correction method was developed based on laboratory experiments to correct for the results in the post-processing phase. The correction function was applied to three MAAP datasets from Gual Pahari (India), Beijing (China), and Welgegund (South Africa). In Beijing, the results could also be compared against a photoacoustic spectrometer (PAS). The correction improved the quality of all three MAAP datasets substantially, even though the individual instruments operated at different flow rates and in different environments.

1 Introduction

A widely used method for measuring atmospheric black carbon (BC) mass concentration involves the determination of absorption of an aerosol sample collected on an appropriate filter matrix. The most common instruments utilized

today for this purpose are the filter-tape-based Aethalometer (Hansen et al., 1984), Multi-Angle Absorption Photometer (MAAP) (Petzold et al., 2002; Petzold and Schönlinner, 2004), and the single-filter-based Particle Soot Absorption Photometer (PSAP) (e.g., Bond et al., 1999). Since BC by definition cannot be unambiguously measured with these instruments, it is customary to refer to the measured carbonaceous light absorbing aerosol constituent as equivalent BC (BC_e) or light-absorbing carbon (LAC). For the sake of simplicity, we use the term BC throughout. For a detailed discussion of the nomenclature used for black carbon or light-absorbing carbon components of the atmospheric aerosol, see, e.g., Bond and Bergstrom (2006) and Andreae and Gelencsér (2006).

It is well known that filter-based BC measurements suffer from several artifacts. These include the filter loading effect that causes a decrease in the measured BC concentration with increasing filter load, and the sample matrix effect that causes scattering aerosols on the filter to increase the measured BC concentration. These artifacts can be corrected to some extent by using different numerical methods (e.g., Bond et al., 1999; Weingartner et al., 2003; Arnott et al., 2005; Virkkula et al., 2007; Collaud Coen et al., 2010). All of the correction schemes have their advantages and disadvantages under field conditions. Thus far, the MAAP has been deemed as the most reliable filter-based instrument for measurement of BC, since the instrument design and software take the typical filter-related artifact effects into account.

We have conducted aerosol field measurements in Gual Pahari (India) from December 2007 to January 2010, including BC measurements with the MAAP (Hyvärinen et al., 2010). During this campaign, we observed that at high BC concentrations the MAAP is not free of measurement artifacts. The observed artifact is different from those seen with other filter-based BC instruments, and to our knowledge has not been reported before in the literature. Here, we quantify this artifact with the assistance of laboratory measurements utilizing two MAAPs operating at different flow rates. The focus of this paper is to raise awareness of the MAAP artifact within the aerosol community, and to demonstrate how the artifact can be circumvented by logging the reflected photodetector signal. In addition, we present a method for correcting the results from the typical instrument print formats in the post-processing phase. The correction is applied to three MAAP datasets: Gual Pahari (India) (Hyvärinen et al., 2010); Beijing (Garland et al., 2009) (China), and Welgegund (South Africa) (Beukes et al., 2012; www.welgegund.org). In Beijing, the results could be compared against a photoacoustic spectrometer (PAS; Garland et al., 2009).

2 Multi-Angle Absorption Photometer

The Thermo Scientific Model 5012 MAAP measures the aerosol BC mass concentration at a single nominal




Fig. 1. Schematic of the MAAP. The transmitted light is measured with the photodetector at $\theta_0 = 0^\circ$, and the reflected light with the photodetectors at $\theta_1 = 130^\circ$ and $\theta_2 = 165^\circ$. (Figure from: “Aerosol Science and Technology: Evaluation of Multiangle Absorption Photometry for Measuring Aerosol Light Absorption”, 39, 40–51, Copyright 2005, Mount Laurel, NJ, reprinted with permission.)

wavelength 670 nm. However the true wavelength has later been measured to be 637 nm (Müller et al., 2011). The typical filter-loading-related artifacts are already taken into account in the design and the internal programming of the instrument. In inter-comparison tests, the MAAP has been found to give reliable results of light absorption by aerosols (e.g., Sheridan et al., 2005; Petzold et al., 2005).

The principle of measuring the absorption coefficient (b_{AP}) using multi-angle absorption photometry has been well documented during instrument development (Petzold and Schönlinner, 2004). The key principle is that, in addition to the typical transmission measurement, the signals scattered to angles at 130 and 165° are also measured (Fig. 1). Additionally, radiative processes are modeled by a radiative transfer scheme for the particle-loaded filter, the aerosol-filter layer alone, and the blank filter alone (Hänel, 1987). The final output of the radiative model is the single scattering albedo (ω_{FILTER}) and the optical depth (τ_{FILTER}) of the aerosol loaded filter layer that match the measured transmitted and reflected signals. From these values, the absorption coefficient of the MAAP ($b_{AP,MAAP}$) is given by the following equation (Petzold et al., 2005):

$$b_{AP,MAAP} \approx b_{ATN,MAAP} = -\frac{A}{V} (1 - \omega_{\text{FILTER}}) \times \tau_{\text{FILTER}} \quad (1a)$$

where $b_{ATN,MAAP}$ is a method-dependent coefficient related to absorption, A is the filter spot area, and V is the sampled volume.

Since the values of ω_{FILTER} and τ_{FILTER} at a given time t always refer to the initial values of the particle-free filter, the incremental increase in $b_{AP,MAAP}$ during time interval $t_i - t_{i-1}$ is determined from

$$b_{AP,MAAP}(t_i) = -\frac{A}{V} \times [(1 - \omega_{\text{FILTER}}(t_i)) \times \tau_{\text{FILTER}}(t_i) - (1 - \omega_{\text{FILTER}}(t_{i-1})) \times \tau_{\text{FILTER}}(t_{i-1})] \quad (1b)$$

The air flow is drawn through a glass fiber filter tape, and the ambient aerosol is collected on a sampling spot of $A = 2 \text{ cm}^2$

area. The sample volume flow through the instrument is measured continuously by the pressure drop across an orifice. For default instrument settings, the filter tape is moved forward to the next blank sampling spot after the transmission of the particle-loaded spot has decreased below 20 %. The initial signals at detection angles 0, 130 and 165° for the particle-free sample spot are determined after the filter spot change during a zeroing procedure, while respective values for the particle-loaded filter are measured at distinct time intervals during aerosol sampling.

In the commercial software version provided by Thermo Instruments, the attenuation of light by the deposited aerosol is measured in time steps of 1 min during continuous aerosol sampling. Values referring to longer time intervals are calculated as averages from the basic 1 min data.

In addition to the MAAP method that utilizes the instrument's internal algorithm, the absorption coefficient can also be determined from the photodetector raw signals in a post-processing procedure. The typically used print formats of the MAAP do not give the raw signals as an output, and have to be logged, e.g., by using the scientific print format 12. These raw signals include the photodetector response signal at 0° (transmittance), and signals at 130 and 165° (reflectance). From the 0 and 165° signals, the light attenuation (ATN) by the sample can be determined as

$$b_{\text{ATN,TRANS}} = \frac{A}{V} \ln \left(\frac{T_0}{T} \right) \quad (2)$$

$$b_{\text{ATN,REFL}} = 0.5 \times \frac{A}{V} \ln \left(\frac{R_0}{R} \right) \quad (3)$$

where (T_0/T) and (R_0/R) are the ratios of photodetector signals at 0 and 165° for a particle loaded and a particle-free filter, respectively. The factor 0.5 in the reflectance measurement results from the fact that the light passes through the layer of sampled aerosol twice before reaching the photodetector.

The measured properties $b_{\text{ATN,TRANS}}$ and $b_{\text{ATN,REFL}}$ cannot be used directly to obtain the BC mass concentration, because typical filter loading artifacts affect the measured signals. Petzold et al. (2005) determined relationships for the b_{AP} and $b_{\text{ATN,TRANS}}$ as well as $b_{\text{ATN,REFL}}$ by utilizing test aerosols. These test aerosols consisted of pure black aerosol samples from kerosene flame particles, and externally mixed gray and black aerosols of varying single scattering albedo. The correction functions are

$$b_{\text{AP,TRANS}} = b_{\text{ATN,TRANS}} \times (0.654 + 3.314 T/T_0)^{-1} \times (1.0 + 0.0015 \exp(\omega_0/0.17))^{-1} \quad (4)$$

$$b_{\text{AP,REFL}} = b_{\text{ATN,REFL}} \times (0.226 + 1.415 R/R_0)^{-1} \quad (5)$$

where ω_0 is the single scattering albedo of the aerosol.

Finally, the absorption coefficients obtained with different methods (MAAP, TRANS and REFL) are directly proportional to the BC mass concentration ($\text{BC}_{\text{METHOD}}$) by a factor

of $1/\sigma_{\text{BC}}$, where $\sigma_{\text{BC}} = 6.6 \text{ m}^2 \text{ g}^{-1}$, the mass-specific absorption cross section of BC at a wavelength of 637 nm.

The increase of BC mass deposited on the filter spot during one time interval of sampling, or BC mass accumulation rate ΔBC in $\mu\text{g cm}^{-2}$, is calculated accordingly as

$$\Delta \text{BC}(t_i) = [(1 - \omega_{\text{FILTER}}(t_i)) \times \tau_{\text{FILTER}}(t_i) - (1 - \omega_{\text{FILTER}}(t_{i-1})) \times \tau_{\text{FILTER}}(t_{i-1})] \times \frac{1}{\sigma_{\text{BC}}} \quad (6)$$

The value of σ_{BC} was determined from a series of lab and ambient measurements (Petzold and Schönlinner, 2004) against gravimetric and thermal reference methods. It is in close agreement with the value of $7.5 \pm 1.2 \text{ m}^2 \text{ g}^{-1}$ at 550 nm for “fresh” carbonaceous aerosol (Bond and Bergstrom, 2006) taking into account scaling of σ_{BC} with the inverse wavelength.

3 MAAP artifact and correction

3.1 Initial observations

The artifact was first noticed during measurements at the EU-CAARI station in Gual Pahari (India). The station was located about 25 km south of New Delhi. Typical for the Indo-Gangetic plains, the region is heavily polluted. For the measurement campaign period of December 2007–January 2010, the PM_{10} and $\text{PM}_{2.5}$ average values were $216 \mu\text{g m}^{-3}$ and $126 \mu\text{g m}^{-3}$, respectively (Hyvärinen et al., 2010). The average BC mass concentration was found to be $12.3 \mu\text{g m}^{-3}$. The BC measurements were conducted utilizing a PM_{10} inlet, and the aerosol was dried by a diffusion drier prior to entering the instrument. The MAAP (Thermo Scientific Model 5012) was run at a nominal flow rate of 8 L min^{-1} . The data were mostly saved as 5 or 1 min averages.

During high concentration periods, an artificial decrease in concentration was observed in the MAAP data at Gual Pahari (Fig. 2). The artifact is related to the filter spot change conducted by the instrument; i.e., when a new blank filter spot is moved into the sampling head, the observed BC concentration decreases. The artifact is clearly not related to the typical filter loading effects, which result in concentration increase rather than decrease after the filter spot change (e.g., Virkkula et al., 2007; Petzold et al., 2005).

For a better understanding of the artifact, we analyzed the instrument raw signals from the MAAP at Gual Pahari. Using Eqs. (1)–(5) and the absorption cross section σ_{BC} , the whole dataset was converted to three different BC values: BC_{MAAP} , BC_{TRANS} and BC_{REFL} . The ω_0 in Eq. (4) for calculating $b_{\text{AP,TRANS}}$ was estimated to be 0.9, which is typical for ambient aerosols. This choice was found to affect the magnitude but not the shape of the transmitted signal.

When comparing the results obtained with the three different methods (Fig. 2), it becomes apparent that the artifact is already present in the raw photodetector signals, especially




Fig. 2. Comparison of BC measured with the MAAP by multi-angle photometry BC_{MAAP} , transmission BC_{TRANS} and reflectance BC_{REFL} methods during a high concentration episode in Gual Pahari, India, 12 August 2009.

in the transmitted 0° signal. BC_{TRANS} almost mimics the behavior of BC_{MAAP} , although showing somewhat higher concentrations in general. BC_{REFL} exhibits a much weaker, yet visible artifact effect. It is also notable in Fig. 2 that, during the few minutes when the filter spot change takes place, there is a strong scattering in the raw signal, which is related to stabilization of the instrument. The exact mechanism causing the artifact is not known, but it seems to be related to erroneous dark counts in the photodetector measuring the transmitted light during the filter zeroing procedure, in combination with an instrument internal averaging procedure of the photodetector raw signals.

3.2 Quantification of the artifact

A laboratory test was conducted to accurately quantify the artifact. The basic assumption for designing the quantification experiment was that the observed artifact is closely related to the BC mass accumulation rate ΔBC in $\mu\text{g min}^{-1}$. Hence, two MAAP instruments were set up to sample from the same aerosol but operated on different volume flow rates. This in turn resulted in different values for ΔBC .

Test aerosol was produced by atomizing a water solution of “Aquadaq”, a soot reference standard (Baumgardner et al., 2012), into a mixing volume of ~ 5 L. Two MAAP instruments sampled from this mixing volume – one (s/n 145) with a high flow rate ($16\text{--}20\text{ L min}^{-1}$), and the other (s/n 87) with a low flow rate ($7\text{--}10\text{ L min}^{-1}$). Make-up air was taken from the lab through a HEPA filter. The concentration of BC was changed by changing the flow rate of the atomizer. MAAP flow rates were calibrated against a Gilibrator bubble flow meter. The concentration- and flow rate-ratios were controlled so that the artifact occurred only in the high flow rate

MAAP, thus utilizing the low flow rate MAAP as a reference instrument.

As the photodetector response (and thus the mass accumulation rate) depends both on the BC mass concentration and the flow rate Q of the instrument, we express the laboratory results in terms of $\Delta BC = BC \times Q$, i.e., accumulation rates. We prefer to use this notation, as both BC and Q can be logged with the MAAP standard print formats. Cumulatively, this also becomes the mass on the filter spot, m . The results indicate that at BC mass accumulation rates $> \sim 0.04\text{ }\mu\text{g min}^{-1}$ the artifact can be identified from the data. The observed artifact is very systematic. After an initial drop, the BC signal increases with a rate proportional to the prevailing BC concentration. The artifact can be roughly divided in three distinct regions (Fig. 3a):

1. At moderately high mass accumulation rates ($0.04\text{ }\mu\text{g min}^{-1} < BC \times Q < 0.08\text{ }\mu\text{g min}^{-1}$), only the very first minutes experience a decrease of the signal, followed by a prominent increase above the initial signal level, before stabilization back to the correct level occurs.
2. At high mass accumulation rates ($0.08\text{ }\mu\text{g min}^{-1} < BC \times Q < 0.14\text{ }\mu\text{g min}^{-1}$), the initial decrease of the signal becomes more apparent compared to the following overestimation. However, the point where the concentration stabilizes back to the correct level is still very distinct, as it is characterized by 1 or 2 min of clearly higher concentrations.
3. At very high mass accumulation rates ($BC \times Q > 0.14\text{ }\mu\text{g min}^{-1}$), the initial signal decrease is so strong that the signal never recovers to the correct level before the next filter spot change, leading to an inevitable underestimation of the BC concentration. This reveals that an erroneous temporal response is not the only outcome of the artifact. At high enough BC concentrations, the MAAP underestimates BC values entirely (Fig. 3b).

3.3 Correction algorithm

To compile a correction algorithm, we addressed the two principal problems present in the original data, i.e., (1) the overall concentrations, which are underestimated when the rising MAAP signal cannot reach the true concentration (Fig. 3b), and (2) the temporal response of the erroneous concentrations (Fig. 3a).

1. The overall correction was made by correlating the smoothed data from the high flow rate MAAP with the data from the reference low flow rate MAAP. The smoothed concentrations, BC_{smooth} , could simply be the last few minutes before the next filter spot change, by removing the other data suffering from the artifact. The correlation can be described with a third level polynomial resulting in the corrected $BC_{\text{CORR}} \times Q$:




Fig. 3. (a) Examples of the artifact temporal response. Symbol lines indicate data from the high flow rate MAAP, and solid lines data from the reference MAAP. Regimes 1, 2 and 3 correspond to accumulation rates ($BC \times Q$) of $0.04\text{--}0.08 \mu\text{g min}^{-1}$, $0.08\text{--}0.14 \mu\text{g min}^{-1}$ and $> 0.14 \mu\text{g min}^{-1}$, respectively. See text for details. (b) Comparison of the low flow rate reference (BC_{ref}) and high flow rate (BC_{MAAP}) BC concentrations.

$$BC_{\text{CORR}} \times Q = 5.665 \pm 0.25 (BC_{\text{smooth}} \times Q)^3 + 0.203 \pm 0.113 (BC_{\text{smooth}} \times Q)^2 + 0.9363 \pm 0.0116 (BC_{\text{smooth}} \times Q). \quad (7)$$

In conditions with changing concentrations, only considering data from a few minutes before each spot change might be misleading due to the poor time resolution.

2. In order to smooth the temporal response, we assumed that the real concentration is a sum of the extracted artifact signal and real changes in the concentrations, and may thus be expressed as

$$BC_{\text{smooth}} = BC_{\text{ini}} + (BC_{\text{meas}} - BC_{\text{artifact}}) \quad (8)$$

where BC_{ini} is the concentration before the spot change, BC_{meas} the measured non-corrected concentration and BC_{artifact} the artifact signal dependent on the initial concentration described below.

The shape of the MAAP artifact signal can be described with the so-called Hill function as a function of the mass of BC on the filter spot, m :

$$BC_{\text{artifact}} \times Q = BC_{\text{max}} \times Q \frac{m^n}{m^n + k^n} \quad (9)$$

where $BC_{\text{max}} \times Q$ is the maximum plateau value simulated by the Hill function, and k and n are parameters describing the slope of the rising mass accumulation rate.

The measured accumulated mass was chosen as the base for characterizing the artifact, as it is a reasonable assumption that the artifact is dependent on both the initial mass accumulation rate and the change of accumulated mass on the filter spot. The laboratory cases were fitted with this function, and the parameters $BC_{\text{max}} \times Q$, k and n were optimized (Supplement). The parameters can be expressed with the following functions and constants:

$$BC_{\text{max}} \times Q = 0.8792 \times (BC_{\text{ini}} \times Q) + 0.0347$$

$$k = 1.6623 \times (BC_{\text{ini}} \times Q) + 0.0462$$

$$n = 20.02 \times (BC_{\text{ini}} \times Q)^2 - 4.6454 \times (BC_{\text{ini}} \times Q) + 1.428$$

where $BC_{\text{ini}} \times Q$ is the mass accumulation rate in $\mu\text{g min}^{-1}$ before the spot change. As noted earlier, when $BC \times Q < 0.14 \mu\text{g min}^{-1}$, the artifact signal eventually makes an abrupt decrease back to the correct level (see Fig. 3a). This point was found to follow the relation:

$$m_d = 0.1632 \times \exp(21.798 \times (BC_{\text{ini}} \times Q)) - 0.4. \quad (10)$$

The algorithm is not able to predict this decrease back to the real signal level, and should not be applied if the accumulated mass from the spot change is greater than m_d (in μg).

An overall representation of the laboratory results is illustrated in Fig. 4. The smoothed data (BC_{smooth}) correct for the temporal response, but not the overall underestimation. However, the polynomial regression (Eq. 7) brings the two datasets to a 1 : 1 ratio ($R^2 = 0.99$). Equation (7) is valid for the mass accumulation rate ($BC_{\text{smooth}} \times Q$) range of $0\text{--}0.39 \mu\text{g min}^{-1}$. The equation is nearly linear in the range of




Fig. 4. High flow rate MAAP BC concentrations as a function of the reference (low flow rate) MAAP concentrations. BC_{MAAP} is the original signal; BC_{smooth} is the smoothed signal without an overall concentration correction. BC_{CORR} is the smoothed signal with the overall concentration correction. BC_{REFL} is the original signal determined with the reflectance method. The 1 : 1 ratio line is also shown.

$0\text{--}0.14\text{ }\mu\text{g min}^{-1}$ with an average relative difference of 3.8 % from a linear 1 : 1 correlation. The upper limit here is restricted by the artifact arising in the reference MAAP. The low flow rate data were used up to $BC \times Q = 0.14\text{ }\mu\text{g min}^{-1}$ so that only the artifact-free data were chosen for the correlation. The average absolute deviation of the corrected measurement points of the high flow rate MAAP compared to the low flow rate MAAP was $0.49\text{ }\mu\text{g m}^{-3}$, and the corresponding average relative deviation was 5.4 %. The highest individual deviations typically occur during the first 2–3 min after the filter spot change, due to the very steep signal increase of the artifact.

3.4 Correction from the raw reflectance signal

Similarly to field observations in Gual Pahari, the laboratory dataset was also converted to BC_{REFL} using Eqs. (3) and (5). We see that these data follow the reference concentrations closely (Fig. 4), although with more scatter ($R^2 = 0.96$). Obviously the reflectance signal is not affected by the measurement artifact. This observation may indicate that the measurement artifact is occurring only in the processed transmission signal, which then affects also the final MAAP output signal while the reflectance signal only, with appropriately applied corrections of the filter matrix effect (Petzold et al., 2005), is not affected by the artifact and reports accurate BC mass concentration values. This presents an opportunity to utilize the reflected signal from the MAAP when mass accumulation rates are high enough that the artifact appears. However, as most of the print formats of the MAAP do not give the raw signals as an output, the algorithm is a useful way for correcting the results.

4 Application of correction to ambient measurements

Three ambient datasets were chosen for testing the correction algorithm: Gual Pahari (India) (Hyvärinen et al., 2010); Beijing (Garland et al., 2009) (China), and Welgegund (South Africa) (Beukes et al., 2012; www.welgegund.org). All these locations suffer from such high BC concentrations that the measurement artifact could be observed from the data.

In Gual Pahari, the MAAP was run at 8 L min^{-1} . The correction algorithm was applied to the full dataset from 14 December 2007 to 19 January 2010. In Beijing, the flow rate of the MAAP was measured to be 9.2 L min^{-1} . The algorithm was applied to a short-term dataset from 10 August 2006 to 9 September 2006, for which we could additionally utilize a PAS as reference method for the absorption of BC. Finally, in Welgegund (South Africa) a MAAP was run at a flow rate of 16.7 L min^{-1} . These data covered the period from 1 June 2010 to 31 August 2010.

Using the algorithm (Eqs. 7–10) clearly improves the temporal response of the MAAP signal, removing most of the signal decreases observed in the uncorrected data from all three locations (Fig. 5a, c and e). On occasion, the first points after a filter spot change still show a concentration decrease for BC_{CORR} , which is due to the steep shape of the artifact: a small uncertainty in m (accumulated mass) can lead to a large uncertainty in BC. In addition, problematic situations may occur when the true concentration exhibits a high gradient during the filter spot change. This can happen especially if there are short-term pollution episodes taking place. In such cases the assumption that the last value of BC on the previous filter spot equals the initial concentration is not valid. This has direct consequences on both the modeled new concentration and the length of the artifact effect. If there is a BC concentration decrease during the filter spot change, the concentrations would be underestimated and the last point of the artifact, m_d , would be overestimated. For an increase, the effect is the opposite. As seen in the example figures, BC_{REFL} may occasionally show values lower (Fig. 5a) or higher (Fig. 5e) than BC_{CORR} . However, the overall trends produced by the two methods are very similar.

In order to better evaluate the performance of the suggested correction algorithm, we compared the original BC_{MAAP} and the corrected BC_{CORR} against BC_{REFL} from all three ambient locations (Fig. 5b, d and f). The erroneous temporal response of BC_{MAAP} is again evident from the downward “tails” appearing in the figures. In addition, BC_{MAAP} underestimates the higher concentrations, similarly to what was seen in the laboratory experiments. Linear regressions fitted to the data (Table 1) do not indicate a substantial difference between the relations BC_{MAAP} vs. BC_{REFL} and BC_{CORR} vs. BC_{REFL} . However, this is mostly because the median concentration is well below $10\text{ }\mu\text{g m}^{-3}$ at all the locations. The regressions do reveal that the correlation at all locations is improved by the correction. For Welgegund, the correlation is worst ($R^2 = 0.94$),




Fig. 5. Examples of the correction algorithm applied to the datasets from Gual Pahari (a, b), Beijing (c, d), and Welgegund (e, f). In panels (b), (d), and (f) the datasets from different methods are compared against the reflected BC signal, BC_{REFL} . Lines in these figures are linear fits to data. BC in Beijing was measured with 1 : 1 dilution. For Beijing, the absorption coefficients b_{METHOD} at 532 nm from the PAS and those derived from the MAAP (see text for details) are also shown.

and in general BC_{REFL} is higher than BC_{CORR} . At this location, high BC concentrations are related to pollution episodes, and 90 % of data points are below $\sim 3 \mu\text{g}/\text{m}^3$. It is possible that the poorer correlation in Welgegund is related to the assumptions made in the determination of

BC_{REFL} . Also in Gual Pahari, BC_{REFL} is generally higher than BC_{CORR} . Opposite to Welgegund, here the difference is probably related to the very high overall concentrations (90th percentile = $24.5 \mu\text{g}/\text{m}^3$, maximum $73.1 \mu\text{g}/\text{m}^3$). At mass accumulation rates higher than the applicability of the

Table 1. Linear regression analysis of BC_{MAAP} , BC_{CORR} and BC_{PAS} against BC_{REFL} at three ambient locations. 10th percentile, median and 90th percentile of the concentrations (in $\mu\text{g m}^{-3}$) are also shown.

Measurement site	Data points	10th perc.	Median	90th perc.	Method	<i>A</i>	<i>B</i>	<i>R</i> ²
Gual Pahari	~ 240 000	2.3	7.6	24.5	MAAP	0.73	2.10	0.91
					CORR	0.92	0.84	0.96
Beijing	~ 41 000	1.1	4.2	10.1	MAAP	0.94	2.08	0.98
					CORR	1.00	−1.20	0.99
					PAS	1.14	1.08	0.97
Welgegund	~ 130 000	0.3	1.1	2.8	MAAP	0.75	0.29	0.86
					CORR	0.90	0.05	0.94

$$BC_{\text{METHOD}} = A * BC_{\text{REFL}} + B$$

correction function ($0.4 \mu\text{g min}^{-1}$), concentrations may be underestimated. While we cannot say with certainty which method produces the most accurate results, it has to be kept in mind that BC_{REFL} is corrected from the raw photomultiplier signal by an empirical function based on test aerosols (Eq. 5). In ambient conditions, especially of high loading with strongly scattering aerosol, Eq. (5) may not be valid. The results do, however, indicate that the artifact correction based on the laboratory experiment may be applied to different ambient environments.

Finally, we were able to compare the results from a PAS against those derived from the MAAP in Beijing. The absorption coefficient values from MAAP at 637 nm were converted to those at 532 nm (PAS wavelength) by assuming an absorption Ångström exponent of 1. In addition, the MAAP in Beijing was run with 1 : 1 dilution, while the PAS sampled without dilution, so the MAAP absorption coefficients were further multiplied by 2. Although the results from the MAAP are slightly lower (by 15 %) than those reported from the PAS (Fig. 5d and Table 1), the difference can be considered acceptable for a correction scheme. Similar and consistent results compared to a PAS were obtained, when the algorithm was applied to MAAP data from another megacity region, Guangzhou, China (Garland et al., 2008; R. M. Garland, private communication, 2012). This agreement further confirms our laboratory findings.

5 Conclusions

We have observed a measurement artifact in the MAAP at high BC concentrations. The artifact is related to the filter spot change – as mass is accumulated on a fresh filter spot, the photodetector response of the transmitted 0° light is lower than anticipated. However, the 165° photodetector signal is not compromised. The artifact seems to be related to erroneous dark counts in the transmitted light photodetector, in combination with an instrument internal averaging procedure of the photodetector raw signals. The artifact behavior however appears to be entirely related to the currently

implemented data inversion algorithm, but not to any unknown physical processes. Using raw data on a 1 Hz basis and post-processing the data independently by an algorithm similar to that described by Petzold and Schönlinner (2004) shows no artifacts as described here (T. Onasch, private communication, 2011). The artifact can be observed if the BC mass accumulation rate $BC \times Q$ exceeds $0.04 \mu\text{g min}^{-1}$. At the typical flow rate of $1 \text{ m}^3 \text{ h}^{-1}$, this relates to a BC concentration of $\sim 3 \mu\text{g m}^{-3}$. Overall concentrations of uncorrected MAAP data are underestimated if $BC \times Q$ exceeds $0.14 \mu\text{g min}^{-1}$. With increasing BC accumulation rate, the underestimation may be several tens of percent.

We compiled an algorithm to correct the BC estimation from the typically most commonly used print formats of the MAAP. The algorithm is not dependent on the saving interval of the data and takes the instrument flow rate into account. The algorithm was tested on data originating from three different ambient environments, and was found to improve all the datasets considerably. In principle, the artifact can also be avoided by diluting the sampled air, but this will result in a loss of accuracy at lower concentrations. The MAAP-reflected signal may also be used to derive correct concentration levels. Therefore, it is strongly recommended to log the raw reflectance signals of MAAPs in highly polluted environments. However, utilizing solely the reflected signal may result in an increased noise in the data.

An updated version of the MAAP firmware is currently in preparation for distribution. However, the correction algorithm as described here is urgently needed for correcting data from worldwide operated MAAP instruments.

Supplementary material related to this article is available online at: <http://www.atmos-meas-tech.net/6/81/2013/amt-6-81-2013-supplement.pdf>.

Acknowledgements. EUCAARI and The Finnish Foreign Ministry's "Particulate pollution and the Indian Brown cloud associated with it" are acknowledged for funding the measurements in India.

Academy of Finland: 132 640, Atmospheric monitoring capacity building in Southern Africa, [2010–2012]

Academy of Finland: 117 505, Air pollution in Southern Africa (APSA) [2006–2009]

Nesslingin säätiö: Ilmansaasteiden ja lumisateen vuorovaikutus arktisilla alueilla

EU LIFE+ project LIFE09 ENV/FI/000572 MACEB

The contributions of R. M. Garland, M. O. Andreae, and U. Pöschl were supported by the Max Planck Society, Germany.

We thank Aki Virkkula for the helpful discussions. We also thank Timothy B. Onasch from Aerodyne Research Inc. for in-depth discussions on the observed measurement artifact and potential explanations.

Edited by: W. Maenhaut

References

- Andreae, M. O. and Gelencsér, A.: Black carbon or brown carbon? The nature of light-absorbing carbonaceous aerosols, *Atmos. Chem. Phys.*, 6, 3131–3148, doi:10.5194/acp-6-3131-2006, 2006.
- Arnott, W. P., Hamasha, K., Moosmüller, H., Sheridan, P. J., and Ogren, J. A.: Towards Aerosol Light-Absorption Measurements with a 7-Wavelength Aethalometer: Evaluation with a Photoacoustic Instrument and 3-Wavelength Nephelometer, *Aerosol. Sci. Tech.*, 39, 17–29, 2005.
- Baumgardner, D., Popovicheva, O., Allan, J., Bernardoni, V., Cao, J., Cavalli, F., Cozic, J., Diapouli, E., Eleftheriadis, K., Genberg, P. J., Gonzalez, C., Gysel, M., John, A., Kirchstetter, T. W., Kuhlbusch, T. A. J., Laborde, M., Lack, D., Müller, T., Niessner, R., Petzold, A., Piazzalunga, A., Putaud, J. P., Schwarz, J., Sheridan, P., Subramanian, R., Swietlicki, E., Valli, G., Vecchi, R., and Viana, M.: Soot reference materials for instrument calibration and intercomparisons: a workshop summary with recommendations, *Atmos. Meas. Tech.*, 5, 1869–1887, doi:10.5194/amt-5-1869-2012, 2012.
- Beukes, J. P., Vakkari, V., van Zyl, P. G., Venter, A. D., Josipovic, M., Jaars, K., Tiitta, P., Kulmala, M., Worsnop, D., Pienaar, J. J., Järvinen, E., Chellapermal, R., Ignatius, K., Maalick, Z., Cesnulyte, V., Ripamonti, G., Laban, T. L., Skrabalova, L., du Toit, M., Virkkula, A., and Laakso, L.: Source region plume characterisation of the interior of South Africa, as measured at Welgegund, in preparation, 2012.
- Bond, T. C. and Bergstrom, R. W.: Light Absorption by Carbonaceous Particles: An Investigative Review, *Aerosol Sci. Tech.*, 40, 27–67, 2006.
- Bond, T. C., Anderson, T. L., and Campbell, D.: Calibration and intercomparison of filter-based measurements of visible light absorption by aerosols, *Aerosol Sci. Tech.*, 30, 582–600, 1999.
- Collaud Coen, M., Weingartner, E., Apituley, A., Ceburnis, D., Fierz-Schmidhauser, R., Flentje, H., Henzing, J. S., Jennings, S. G., Moerman, M., Petzold, A., Schmid, O., and Baltensperger, U.: Minimizing light absorption measurement artifacts of the Aethalometer: evaluation of five correction algorithms, *Atmos. Meas. Tech.*, 3, 457–474, doi:10.5194/amt-3-457-2010, 2010.
- Garland, R. M., Yang, H., Schmid, O., Rose, D., Nowak, A., Achtert, P., Wiedensohler, A., Takegawa, N., Kita, K., Miyazaki, Y., Kondo, Y., Hu, M., Shao, M., Zeng, L. M., Zhang, Y. H., Andreae, M. O., and Pöschl, U.: Aerosol optical properties in a rural environment near the mega-city Guangzhou, China: implications for regional air pollution, radiative forcing and remote sensing, *Atmos. Chem. Phys.*, 8, 5161–5186, doi:10.5194/acp-8-5161-2008, 2008.
- Garland, R. M., Schmid, O., Nowak, A., Achtert, P., Wiedensohler, A., Gunthe, S. S., Takegawa, N., Kita, K., Kondo, Y., Hu, M., Shao, M., Zeng, L. M., Zhu, T., Andreae, M. O., and Pöschl, U.: Aerosol optical properties observed during CAREBeijing-2006: Characteristic differences between the inflow and outflow of Beijing city air, *J. Geophys. Res.-Atmos.*, 114, D00G04, doi:10.1029/2008JD010780, 2009.
- Hänel, G.: Radiation budget of the boundary layer: Part II. Simultaneous measurement of mean solar volume absorption and extinction coefficients of particles, *Beitr. Phys. Atmos.*, 60, 241–247, 1987.
- Hansen, A. D. A., Rosen, H., and Novakov, T.: The aethalometer – an instrument for the real-time measurement of optical absorption by aerosol particles, *Sci. Total Environ.*, 36, 191–196, 1984.
- Hyvärinen, A.-P., Lihavainen, H., Komppula, M., Panwar, T. S., Sharma, V. P., Hooda, R. K., and Viisanen, Y.: Aerosol measurements at the Gual Pahari EUCAARI station: preliminary results from in-situ measurements, *Atmos. Chem. Phys.*, 10, 7241–7252, doi:10.5194/acp-10-7241-2010, 2010.
- Müller, T., Henzing, J. S., de Leeuw, G., Wiedensohler, A., Alastuey, A., Angelov, H., Bizjak, M., Collaud Coen, M., Engström, J. E., Gruening, C., Hillamo, R., Hoffer, A., Imre, K., Ivanow, P., Jennings, G., Sun, J. Y., Kalivitis, N., Karlsson, H., Komppula, M., Laj, P., Li, S.-M., Lunder, C., Marinoni, A., Martins dos Santos, S., Moerman, M., Nowak, A., Ogren, J. A., Petzold, A., Pichon, J. M., Rodriguez, S., Sharma, S., Sheridan, P. J., Teinilä, K., Tuch, T., Viana, M., Virkkula, A., Weingartner, E., Wilhelm, R., and Wang, Y. Q.: Characterization and intercomparison of aerosol absorption photometers: result of two intercomparison workshops, *Atmos. Meas. Tech.*, 4, 245–268, doi:10.5194/amt-4-245-2011, 2011.
- Petzold, A. and Schönlinner, M.: Multi-Angle Absorption Photometry – A New Method for the Measurement of Aerosol Light Absorption and Atmospheric Black Carbon, *J. Aerosol Sci.*, 35, 421–441, 2004.
- Petzold, A., Kramer, H., and Schönlinner, M.: Continuous Measurement of Atmospheric Black Carbon using a multi-Angle Absorption Photometer, *Environ. Sci. Poll. Res.*, 4, 78–82, 2002.
- Petzold, A., Schloesser, M., Sheridan, P. J., Arnott, W. P., Ogren, J. A., and Virkkula, A.: Evaluation of Multi-Angle Absorption Photometry for Measuring Aerosol Light Absorption, *Aerosol Sci. Tech.*, 39, 40–51, 2005.

- Sheridan, P. J., Arnott, W. P., Ogren, J. A., Andrews, E., Atkinson, D. B., Covert, D. S., Moosmüller, H., Petzold, A., Schmid, B., Strawa, A. W., Varma, R., and Virkkula, A.: The Reno aerosol optics study: An evaluation of aerosol absorption measurement methods, *Aerosol Sci. Tech.*, 39, 1–16, 2005.
- Virkkula, A., Mäkelä, T., Hillamo, R., Yli-Tuomi, T., Hirsikko, A., Hämeri, K., and Koponen, I. K.: A simple procedure for correcting loading effects of aethalometer data, *J. Air Waste Manage. Assoc.*, 57, 1214–1222, 2007.
- Weingartner, E., Saathoff, H., Schnaiter, M., Streit, N., Bitnar, B., and Baltensperger, U.: Absorption of light by soot particles: determination of the absorption coefficient by means of aethalometers, *J. Aerosol Sci.*, 34, 1445–1463, 2003.